



18 rue Marie et Pierre Curie
92260 Fontenay aux roses
Phone: +33 (0)1 46 60 61 51
Fax: +33 (0)1 78 76 74 18
<http://www.fimel.fr>



Beam Quality Control



- Fits on the linac's accessory shoe
- Large analysis area: 300 x 300 mm
- Resolution: 0.5 mm
- Energy range: ^{60}Co to 25 MV in photons
4 to 25 MeV in electrons
- Real-time image acquisition
- Beam checked using standardized protocols
- Periodic controls automated, AFSSAPS 2007
- MLC control automated
- Projector control automated for HDR brachy
- Data imports/exports
- Lynx/TPS data comparisons
- Dose distribution control, gamma index calculation
- Dimensions: 360 x 370 x 600 mm